

**CERTIFICATE OF ANALYSIS**

SYNERGYCHEM PTE. LTD.  
14 Robinson Road #13-00,  
Far East Finance Building,  
Singapore 048545  
Tel : 65-6240 6870

Fax : 65-6323 1839

Certificate No : CN/nOTS0074/0914  
Sample Log Code : nOTS0061/0914  
Sample Received Date : 23-Sep-2014  
Complete Analysis Date : 14-Oct-2014  
Date Issue : 14-Oct-2014

**Sample description** : Maxicombi AT3300 / AR3303



Photo of sample

**Test performed** : Acute Systemic Toxicity (ISO 10993-11)

**Test procedures:**


- 1) Healthy male mice of inbred strain weighing 17-23 gram were used.
- 2) 50 ml/kg of Maxicombi extracted in saline was injected to a group of 5 mice.
- 3) Do the same as step no. 2 using saline for control.
- 4) The animals were observed for 72 hours after injection.
- 5) During the observation all the animals treated should not show any abnormalities or death.

**Analysis results:**

| Control |                            |                    |          |          | Test Group (Maxicombi AT3300/AR3303 extract) |                            |                    |          |          |
|---------|----------------------------|--------------------|----------|----------|--|----------------------------|--------------------|----------|----------|
| Mice    | Body weight of mice (gram) | Observation period |          |          | Mice   | Body weight of mice (gram) | Observation period |          |          |
|         |                            | 24 hours           | 48 hours | 72 hours |  |                            | 24 hours           | 48 hours | 72 hours |
| 1       | 22                         | no death           | no death | no death | 1  | 23                         | no death           | no death | no death |
| 2       | 22                         | no death           | no death | no death | 2  | 22                         | no death           | no death | no death |
| 3       | 23                         | no death           | no death | no death | 3  | 21                         | no death           | no death | no death |
| 4       | 21                         | no death           | no death | no death | 4  | 20                         | no death           | no death | no death |
| 5       | 21                         | no death           | no death | no death | 5  | 22                         | no death           | no death | no death |

**Conclusion:**

No mice dies or shows any abnormalities within 72 hours. The sample passes the test for Acute Systemic Toxicity.

  
 Teem Chin Mean  
 Manager  
 M.Sc., AMIC,  
 A/2152/4620/04